



# Quantitative Methods in Chemistry

Week 8, Lecture 4

This week: Practice of Chromatography – HPLC, Gas Chromatography, Supercritical Fluid Chromatography, Detectors for analytes

This lecture: Detectors for analytes

NPTTEL




## Requirements from the detectors

- High sensitivity – able to detect nanogram or picogram of analyte is needed
- Selectivity – In certain applications, selective detection of one analyte over many others is needed.
- High signal-to-noise ratio – to clearly identify the presence of analyte in the eluent
- Low baseline noise – to discriminate the presence of analyte
- Fast response – Needed for the online or real-time detection of the analyte during the chromatographic separations
- Low Baseline drift


NPTTEL

## Points to be considered when using detectors



Signal-to-noise ratio

3σ  
 ↑  
 fluctuations in the signal in absence of any analyte




Drift of baseline with time

Possible got in presence of baseline drift  
 →


Stable baseline  
 ↙

- Response Factor = Peak area / Conc. Of Analyte
- Spectral sensitivity of the detectors

## Detector Dynamic Linear Range and Response Time



## Detector Response Time




Where  $\tau$  = time taken by the signal to reach  $1 - (1/e) = 63.2\%$  of the final value.  $e = 2.718$

$$\frac{A \propto C}{A < 1.0}$$



$\tau$  when  $S(t) = 0.632 \times S_f$   
Low  $\tau \Rightarrow$  Rapid detector response

## Detectors in LC

### UV/ vis detectors



### Refractive Index detector or Differential refractometer




$$RI_{ref} \neq RI_{(sample)}$$

Universal detector

$\mu\text{M}$  concn range detection


## Detectors in GC

### Flame Ionization Detector (FID)



- A high temperature flame is used to pyrolyze the analyte molecules eluting out of the column.
- The ions thus generated are collected on an ion collector. This ion current resulting from the pyrolysis of analyte is used as the signal.
- A universal detector for all types of organic compounds. Has high sensitivity of upto ng (ppb levels), large linear dynamic range, and is also less prone to contamination because everything is burnt out!
- It can't be used in SFC with added organic additives to increase the polarity of the mobile phase. It is destructive detector!

## Thermal Conductivity Detector



Filament heated by current flow and cooled by gas flow.

Column eluent containing analyte will have a different cooling efficiency. This may require altering the current flowing through the filament to maintain the temperature.

*Thermal equilibrium unless the eluent contains analyte*

The change in the current flow with regards to the reference cell being purged with only carrier gas is recorded as the signal for this detector.


*$I_{ref} - I_{sample}$   
vs time*

Requirement: Constant flow of the carrier gas and column eluent. Heat conductivity values of the carrier gas should be kept in mind.

Advantages: Universal, Non destructive

Disadvantage: Lower sensitivity than FID (ug or ppm levels)

## Electron Capture Detector



- Uses a mildly radioactive  $Ni^{63}$  foil to cover the inside of a hollow cylinder.
- The  $\beta$ -radiation from  $Ni^{63}$  ionizes the air between the two electrodes.
- The resulting ions are collected on the positive ion detecting electrode.
- The presence of analytes containing halide residues absorb the  $\beta$ -radiation and also quench the ions present in the air.
- $\downarrow$  Baseline signal  $\Rightarrow$   $\uparrow$  analyte containing halide residues.
- Advantages: Most sensitive detector – can detect picogram (or ppt) level analytes.
- Disadvantages: Only for halide containing organics, sensitive to moisture!